

DECLARATION OF CONFORMITY

The manufacturer:

Worldwide Innovations & Technologies, Inc.
14740 W. 101st Terr.
Lenexa, KS 66215 USA

declares that the new PPE described hereafter.....

RADPAD 9100 No Brainer, x-ray protective shielding for providing protection against secondary radiation for the forehead/brain with attenuation at 90 kVp of 50%, 75%, 90% and 95% and nominal minimum lead equivalences of 0.06 mm, 0.125 mm, 0.250 mm and 0.375 mm based on EN 61331-1:2002 and ASTM F2547-06 test requirements as described in the manufacturer's Technical File entitled "9000 Series PPE Technical File"

is in conformity with the provisions of PPE Directive 89/686/EEC:1989-12-21 and Regulation (EU) 2016/42.

is subject to the procedure set out in Article 11B of PPE Directive EEC 89/686 and Annex VIII (Module D) of Regulation (EU) 2016/425 under the supervision of the approved body:

SGS United Kingdom, Limited
202B, Worle Parkway
Weston-super-Mare
BS22 6WA United Kingdom

July 24, 2018
(Date)

Catherine W. Cadwalader

(Signature of manufacturer)

Catherine W. Cadwalader, President and CEO